

SPECIFICATION

APDSTAG(R) Wako Eluent for Amino Acid Automated Analysis

REQUIREMENT	SPECIFICATION
Appearance	Colorless clear liquid
Absorbance (220nm)	max.0.50
Absorbance (230nm)	max.0.10
Absorbance (240nm)	max.0.02
Absorbance (250~400nm)	max.0.01
pH at 25°C	5.95~6.05
Fluorescence test	to pass test
Concentration (20°C)	24.0~26.0mmol/l