

## SPECIFICATION

APDSTAGR Wako Amino Acids Mixture Standard Solution No.2  
for Amino Acid Automated Analysis

REQUIREMENT	SPECIFICATION
Appearance	Colorless clear liquid
Density (20°C)	measured value
Practical test	to pass test
Taurine	4.500~5.500 $\mu$ mol/mL
L-Threonine	4.500~5.500 $\mu$ mol/mL
L-Serine	4.500~5.500 $\mu$ mol/mL
Glycine	9.00~11.00 $\mu$ mol/mL
L-Alanine	9.00~11.00 $\mu$ mol/mL
L-Valine	9.00~11.00 $\mu$ mol/mL
L-Isoleucine	4.500~5.500 $\mu$ mol/mL
L-Leucine	4.500~5.500 $\mu$ mol/mL
L-Tyrosine	4.500~5.500 $\mu$ mol/mL
L-Phenylalanine	4.500~5.500 $\mu$ mol/mL
L-Ornithine	2.250~2.750 $\mu$ mol/mL
L-Lysine	4.500~5.500 $\mu$ mol/mL
L-Histidine	4.500~5.500 $\mu$ mol/mL
L-Arginine	2.250~2.750 $\mu$ mol/mL
L-Proline	4.500~5.500 $\mu$ mol/mL