

SPECIFICATION

Amino Acids Mixture Standard Solution, Type AN (High Range)

for Amino Acid Automated Analysis

REQUIREMENT	SPECIFICATION
Appearance	Colorless clear liquid
O-Phospho-L-serine	1.19~1.31 μ mol/mL
Taurine	1.19~1.31 μ mol/mL
O-Phosphoethanolamine	1.19~1.31 μ mol/mL
Urea	47.5~52.5 μ mol/mL
L-Aspartic Acid	2.38~2.62 μ mol/mL
L-Threonine	2.38~2.62 μ mol/mL
L-Serine	2.38~2.62 μ mol/mL
L-Glutamic Acid	2.38~2.62 μ mol/mL
Sarcosine	5.94~6.56 μ mol/mL
Glycine	2.38~2.62 μ mol/mL
L-Alanine	2.38~2.62 μ mol/mL
L-Citrulline	2.38~2.62 μ mol/mL
DL-2-Aminobutyric Acid	1.19~1.31 μ mol/mL
L-Valine	2.38~2.62 μ mol/mL
L-Cystine	2.38~2.62 μ mol/mL
L-Methionine	2.38~2.62 μ mol/mL
L-Cystathionine	1.19~1.31 μ mol/mL
L-Isoleucine	2.38~2.62 μ mol/mL
L-Leucine	2.38~2.62 μ mol/mL
L-Tyrosine	2.38~2.62 μ mol/mL
L-Phenylalanine	2.38~2.62 μ mol/mL
β -Alanine	2.38~2.62 μ mol/mL
DL-3-Aminoisobutyric Acid	2.38~2.62 μ mol/mL
L-Hydroxyproline	2.38~2.62 μ mol/mL
L-Proline	2.38~2.62 μ mol/mL