

Code No.03-1568

# SPECIFICATION

## Chloroform Standard Solution (1 mg/mL Methanol Solution) for Water Analysis

REQUIREMENT	SPECIFICATION
Appearance	Colorless clear liquid
Practical test	to pass test
Concentration	0.97~1.03mg/mL