

Code No.03-1619

SPECIFICATION

Chromium Standard Solution (Cr 1000)

REQUIREMENT	SPECIFICATION
Appearance	Yellow clear liquid
Acidity (HNO ₃)	0.09~0.11 mol/L
Concentration (20°C)	990~1010mg/L