

SPECIFICATION

Magnesium Perchlorate
for Elemental Analysis

REQUIREMENT	SPECIFICATION
* Appearance	White, crystals or pillar or shot
* Particle size (300~850 μ m)	min.80%
Assay [Mg(ClO ₄) ₂]	min.83.0%(mass/mass)
Acidity (as HClO ₄)	max.0.005%(mass/mass)
Alkalinity (as MgO)	max.0.01%(mass/mass)

JIS K 8228:2020(*:Additional test performed by Wako but not required by JIS)