

SPECIFICATION

10XPBS(-)
for Cell Culture

REQUIREMENT	SPECIFICATION
Appearance	Colorless clear liquid
pH	7.1~7.7
Osmolality	280~320mOsm/kg
Bacterial endotoxins	max.0.1EU/ml
Sterility test	to pass test
Mycoplasma test	to pass test