

Code No.20-1300

SPECIFICATION

0.2mol/L Tris-HCl Buffer Solution  
for Endotoxin Detection

| REQUIREMENT          | SPECIFICATION          |
|----------------------|------------------------|
| Appearance           | Colorless clear liquid |
| pH at 25°C           | 7.9~8.1                |
| Bacterial endotoxins | to pass test           |

Manufacturer FUJIFILM Healthcare Manufacturing Corporation