

SPECIFICATION

Ultrapure Water
for ICP Analysis

REQUIREMENT	SPECIFICATION
Appearance	Colorless clear liquid
Lithium (Li)	max.100ppt
Sodium (Na)	max.100ppt
Magnesium (Mg)	max.100ppt
Aluminium (Al)	max.100ppt
Potassium (K)	max.100ppt
Calcium (Ca)	max.100ppt
Titanium (Ti)	max.100ppt
Vanadium (V)	max.100ppt
Chromium (Cr)	max.100ppt
Manganese (Mn)	max.100ppt
Iron (Fe)	max.100ppt
Cobalt (Co)	max.100ppt
Nickel (Ni)	max.100ppt
Copper (Cu)	max.100ppt
Zinc (Zn)	max.100ppt
Gallium (Ga)	max.100ppt
Germanium (Ge)	max.100ppt
Arsenic (As)	max.100ppt
Strontium (Sr)	max.100ppt
Zirconium (Zr)	max.100ppt
Molybdenum (Mo)	max.100ppt
Silver (Ag)	max.100ppt
Cadmium (Cd)	max.100ppt
Tin (Sn)	max.100ppt
Antimony (Sb)	max.100ppt
Barium (Ba)	max.100ppt
Tungsten (W)	max.100ppt
Lead (Pb)	max.100ppt

