

## SPECIFICATION

WakosilR 60, 64~210  $\mu\text{m}$   
for Column Chromatography

REQUIREMENT	SPECIFICATION
Appearance	White small granules
Particle diameter (d50)	80~110 $\mu\text{m}$
Particle diameter (d10)	min.60 $\mu\text{m}$
Particle diameter (d90)	max.165 $\mu\text{m}$
Precipitation volume	1.6~2.5mL/g
Loss on drying at 105°C	5.0~10.0%
pH(50g/L aqueous slurry, 25°C)	6.5~7.5
Flowing rate	1.6~2.3mL/min
Separation power	to pass test
Activity	to pass test