

SPECIFICATION

Presep(R)-C Alumina for Sample Pretreatment

| REQUIREMENT | SPECIFICATION |
|---------------------------|---------------------|
| Appearance | to pass test |
| Loss on ignition | max.5.0% |
| Weight of column packing | under investigation |
| Permitted pressure (1) | to pass test |
| Permitted pressure (2) | to pass test |
| Leakage of column packing | to pass test |
| Recovery | 95~105% |
| State of column packing | to pass test |