

Code No.29-3677

SPECIFICATION

PresepR C18(ODS)-II (2g/15mL)  
for Sample Pretreatment

REQUIREMENT	SPECIFICATION
Appearance	to pass test
Height of column packing	1.8~2.2cm
Leakage of column packing	to pass test
Recovery	95~110%