

Code No.29-5120

SPECIFICATION

Limulus ES-II Test Wako 0.015  
for Endotoxin Detection

| REQUIREMENT    | SPECIFICATION     |
|----------------|-------------------|
| Appearance     | to pass test      |
| Sensitivity    | 0.015EU/mL        |
| Potency of CSE | 1000~15000EU/vial |
| Practical test | to pass test      |

Manufacturer FUJIFILM Wako Chemicals U.S.A. Corporation