

Code No.29-5120

# SPECIFICATION

## Limulus ES-II Test Wako 0.015 for Endotoxin Detection

REQUIREMENT	SPECIFICATION
Appearance	to pass test
Sensitivity	0.015EU/mL
Potency of CSE	1000~15000EU/vial
Practical test	to pass test

Manufacturer FUJIFILM Wako Chemicals U.S.A. Corporation