

Code No.29-5360

SPECIFICATION

Limulus Color KY Single Test Wako for Endotoxin Detection

REQUIREMENT	SPECIFICATION
Appearance	to pass test
Potency of CSE	1000~15000EU/vial
Practical test	to pass test

Manufacturer FUJIFILM Healthcare Manufacturing Corporation