

Code No.01-0421

SPECIFICATION

Anthracene
Standard Material

REQUIREMENT	SPECIFICATION
Appearance	to pass test
Solubility in toluene	to pass test
Melting point	216~218°C
Loss on drying over silica gel in vacuum	max.0.2%
Assay (capillary GC)	min.99.5%