

Code No.01-1128

## SPECIFICATION

### Acetone for Pesticide Residue And Polychlorinated Biphenyl Tests

Concentration 300

REQUIREMENT	SPECIFICATION
* Appearance	Colorless clear liquid
* Suitability for pesticide residue analysis (GC-FPD)	to pass test
Assay (CH <sub>3</sub> COCH <sub>3</sub> ) (GC)	min.99.5%(area/area)
Density (20°C)	0.789~0.792g/mL
Water	max.0.3%(mass/mass)
Residue after evaporation	max.2ppm(mass/mass)
Suitability for pesticide residue-PCB analysis	to pass test

JIS K 8040:2020(\*:Additional test performed by Wako but not required by JIS)