

SPECIFICATION

o-Dichlorobenzene

for GPC

REQUIREMENT	SPECIFICATION
Appearance	Colorless clear liquid
Density (20°C)	1.303~1.309g/mL
Refractive index 20°C	1.551~1.552
Absorbance (310nm)	max.0.50
Absorbance (320nm)	max.0.10
Absorbance (350nm)	max.0.05
Absorbance (370nm)	max.0.02
Absorbance (380~400nm)	max.0.01
Suitability for GPC test	to pass test
Water	max.0.02%
Residue after evaporation	max.0.001%
Chloride (Cl)	max.0.001%
Free chlorine (Cl)	max.1ppm
Assay (capillary GC)	min.99.0%