

Code No.06-0645

SPECIFICATION

2nd Fluid for disintegration test, pH 6.8 (X10)

for Disintegration and Dissolution Test

REQUIREMENT	SPECIFICATION
Appearance	Colorless clear liquid
pH(25°C)(1→10)	6.75~6.85