

Code No.07-0475

SPECIFICATION

Glycerol

for Fluorescence Microscope

REQUIREMENT	SPECIFICATION
Appearance	Colorless clear liquid
Absorbance (240nm)	max.1.0
Absorbance (260nm)	max.0.5
Absorbance (280nm)	max.0.3
Absorbance (300~400nm)	max.0.1
Water	max.1.0%
Fluorescence test	to pass test
Assay (capillary GC)	min.99.0%