

Code No.13-0814

# SPECIFICATION

## Methanol

for Boron Determination

REQUIREMENT	SPECIFICATION
Appearance	Colorless clear liquid
Refractive index 20°C	1.327~1.330
Water	max.0.05%
Boron (B)	to pass test
Assay (capillary GC)	min.99.7%