SPECIFICATION

PTH-amino Acids Mobile Phase A (for Gradient Elution)

for SHIMADZU Protein Peptide Sequencer

REQUIREMENT	SPECIFICATION
Appearance	Colorless clear liquid
pH at 20℃	5.1~5.4
Practical test	to pass test

(1 / 1) revised on 2022/09/14