

Code No.19-0962

## SPECIFICATION

### Sodium Ion Standard Solution (Na+:1,000ppm)

for Ion Chromatography

REQUIREMENT	SPECIFICATION
Appearance	Colorless clear liquid
Acidity (as HNO <sub>3</sub> )	0.018~0.022mol/L
Foreign cations	to pass test
Concentration (20°C)	0.99~1.01mg/mL