## **SPECIFICATION**

## Presep(R)-C Florisil(R)

for Sample Pretreatment

REQUIREMENT	SPECIFICATION
Appearance	to pass test
Loss on drying at 180℃	max.3.0%
Weight of column packing	under investigation
Permitted pressure (1)	to pass test
Permitted pressure (2)	to pass test
Leakage of column packing	to pass test
Recovery	95~105%
State of column packing	under investigation

(1 / 1) revised on 2007/06/22