

Code No.29-3483

SPECIFICATION

PresepR RPP-WCX (60mg/3mL) for Sample Pretreatment

REQUIREMENT	SPECIFICATION
Appearance	to pass test
Weight of column packing	55~65mg
Leakage of column packing	to pass test
Flowing rate	min.25mL/min
Recovery	80~110%
Runoff rate	to pass test
Ion-exchange capacity	under investigation