

Code No.29-5310

SPECIFICATION

Limulus Color KY Test Wako
for Endotoxin Detection

REQUIREMENT	SPECIFICATION
Appearance	to pass test
Potency of CSE	1000~15000EU/vial
Practical test	to pass test

Manufacturer FUJIFILM Healthcare Manufacturing Corporation