2024/10/12
Distributed by FUJIFILM Wako Pure Chemical Corporation
Manufactured by FUJIFILM Wako Pure Chemical Corporation

Code No.01-1545

SPECIFICATION

6 Aldehydes-DNPH Mixture Standard Solution (each 0.1 μg Aldehyde/ μL Acetonitrile Solution)

for Stinking Substances Analysis (HPLC)

REQUIREMENT	SPECIFICATION
Appearance	Yellow clear liquid
Practical test	to pass test
Formaldehyde-DNPH	0.665～0.735 μg/ μL
Acetaldehyde-DNPH	0.483～0.535 μg/ μL
Propionaldehyde-DNPH	0.389～0.431 μg/ μL
Isobutyraldehyde-DNPH	0.332～0.368 μg/ μL
Isovaleraldehyde-DNPH	0.293～0.325 μg/ μL
n-Valeraldehyde-DNPH	0.293～0.325 μg/ μL

revised on 2022/04/19