

# SPECIFICATION

## Distilled Water for Amino Acid Sequence Analysis

REQUIREMENT	SPECIFICATION
Appearance	Colorless, clear liquid
Absorbance (200~220nm)	max.0.02
Absorbance (230~400nm)	max.0.01
Chloride	to pass test
Ammonium (NH <sub>4</sub> )	(Cl : max. about0.1 ppm) max.0.03ppm
Substances reducing permanganate	to pass test