

Code No.06-0372

SPECIFICATION

Fluorobenzene Standard

for Air Pollutant Analysis

REQUIREMENT	SPECIFICATION
Appearance	Colorless ~ pale yellowish brown, clear liquid
Density (20°C)	under investigation
Refractive index 20°C	1.463~1.467
Water	max.0.1%
Assay (capillary GC)	min.99.0%