

SPECIFICATION

Mayer's Hematoxylin Solution for Pathology Research

REQUIREMENT	SPECIFICATION
Appearance	to pass test
Absorption maximum (λ max1)	550~560nm
Absorption maximum (λ max2)	430~440nm
Absorption maximum (λ max3)	275~285nm
Absorbance (1→100, λ max1)	0.5~0.8
Absorbance (1→100, λ max2)	0.15~0.25
Absorbance (1→100, λ max3)	0.15~0.25
pH at 25°C	2.0~3.0
Practical test	to pass test