

Code No.13-1408

SPECIFICATION

0.1w/v% Methyl Orange Solution

for pH Analysis

REQUIREMENT	SPECIFICATION
Appearance	to pass test
Absorption maximum (pH4.8)	under investigation
Absorbance (1→200, pH4.8)	under investigation
Visual transition interval(pH)	3.1~4.4 (yellowish red ~ reddish yellow)