

Code No.19-0470

SPECIFICATION

Sulfuric Acid

Wako 1st Grade

REQUIREMENT	SPECIFICATION
Appearance	Colorless clear liquid
Residue after ignition (as sulfate)	max.0.002%
Chloride (Cl)	max.1ppm
Nitrate	to pass test
Copper (Cu)	max.1ppm
Lead (Pb)	max.1ppm
Arsenic (As)	max.0.1ppm
Selenium (Se)	max.1ppm
Iron (Fe)	max.3ppm
Substances reducing permanganate	to pass test
Assay	min.95.0%