

SPECIFICATION

20w/v% Silver Nitrate Solution for COD Measurement

REQUIREMENT	SPECIFICATION
Appearance	Colorless clear liquid
Acidity	to pass test
Substances reducing permanganate (as O)	max.1ppm
Assay (20°C)	18.0~22.0w/v%