

Code No.20-1042

SPECIFICATION

0.1 mol/L Tris-HCl Buffer Solution for Endotoxin Detection

REQUIREMENT	SPECIFICATION
Appearance	Colorless clear liquid
pH at 30°C	7.2~7.4
Bacterial endotoxins	to pass test

Manufacturer FUJIFILM Healthcare Manufacturing Corporation