

Code No.20-2101

# SPECIFICATION

## Tungsten Standard Solution (W 1000) for ICP Analysis

REQUIREMENT	SPECIFICATION
Appearance	Colorless clear liquid
Acidity (as HNO <sub>3</sub> )	4.9~5.1w/w%
Acidity (as HF)	0.4~0.6w/w%
Concentration (20°C)	990~1010mg/L
Trace impurities	to pass test