

Code No.29-3344

## SPECIFICATION

### PresepR PFC-II (60mg/3mL)

for Sample Pretreatment

REQUIREMENT	SPECIFICATION
Appearance	to pass test
Weight of column packing	54~69mg
Flowing rate	min.30mL/min
Recovery	85~125%