

Code No.29-4745

## SPECIFICATION

### PresepR-C C18(ODS) (Short)

for Sample Pretreatment

| REQUIREMENT               | SPECIFICATION       |
|---------------------------|---------------------|
| Appearance                | to pass test        |
| Loss on drying at 85°C    | max.5.0%            |
| Weight of column packing  | under investigation |
| Permitted pressure (1)    | to pass test        |
| Permitted pressure (2)    | to pass test        |
| Leakage of column packing | to pass test        |
| Recovery                  | 90~105%             |
| State of column packing   | to pass test        |