

Code No.29-8150

SPECIFICATION

Limulus ES-II Single Test Wako 0.25
for Endotoxin Detection

| REQUIREMENT | SPECIFICATION |
|---------------------|-------------------|
| Appearance | to pass test |
| Solubility in water | to pass test |
| Sensitivity | 0.25EU/mL |
| Potency of CSE | 1000~15000EU/vial |
| Practical test | to pass test |

Manufacturer FUJIFILM Healthcare Manufacturing Corporation