

[LBIS Mouse Insulin ELISA Kit (H-type)]

Cat #630-10371

Please, read this instruction carefully before use.

This kit is manufactured by FUJIFILM Wako Pure Chemical Corporation. Use only the current version of Instruction Manual enclosed with the kit!

1. Intended use

LBIS Mouse Insulin ELISA Kit (H-type) is a sandwich ELISA system for quantitative measurement of mouse insulin. This kit is intended for research use only.

2. Storage and expiration

When the complete kit is stored at 2 °C - 8 °C (Do not freeze), the kit is stable until the expiration date shown on the label on the container. Opened reagents should be used as soon as possible to avoid loss in optimal assay performance caused by storage environment.

3. Introduction

Insulin is a peptide hormone secreted from B cells of islet of Langerhans in the pancreas with a molecular weight of about 5800 and pI 5.4. It is consisted of 2 chains, A and B. It has 3 disulfide bonds formed between A6 and A11, A7 and B7, and A20 and B19. Insulin exists as a dimer molecule in acidic to neutral solution without Zn ion, and as a hexamer including two Zn ions in neutral solution if Zn ions are present. Main targets of insulin are liver, muscle, and adipose tissue. Insulin actions in these targets are as follows. In the liver, it promotes glycogenesis, protein synthesis, fatty acid synthesis, carbohydrate utilization, and inhibition of gluconeogenesis. In the muscle, it promotes membrane permeability for carbohydrates, amino acids and K ion, glycogenesis, protein synthesis, while inhibits protein degradation. In the adipose tissue, it promotes membrane permeability for glucose and fatty acid synthesis.

A precursor of insulin, called proinsulin with a single polypeptide chain, is first synthesized in the cell, then sulfide bonds are formed, and finally by enzymatic cutting at two sites, active insulin and c-peptide (connecting peptide) are formed. Potency of an insulin preparation was originally determined by bioassay. However, whole body bioassay inevitably shows poor precision owing to individual variation.

WHO issued 1st International Standard for human insulin in 1986 which has the potency of 26 IU/mg (0.038 mg/IU). In the same year, 1st International Standard of bovine insulin, the potency of which is 25.7 IU/mg, and Porcine insulin 1st International Standard, 26 IU/mg, were provided. Before these standards, in 1974, 1st International Reference Preparation of human insulin for immunoassay was provided as 3 IU/ampoule. Based on the above data, if the biological activity of insulin per molecule is the same among various animal species, potencies of animal insulin might be calculated from their molecular weights. But, so far, we do not have experimental proof about this. As the molecular weights of insulin of various animals are nearly the same, and the differences are within 1 %, there may be no critical fault if we think that the general potency of insulin is 26 IU/mg. Rat and mouse have two molecular species of insulin, type 1 and type 2. Amino acid sequences of these molecular species are same between rat and mouse. But as their ratios are different between these two animal species, it is recommended to use standard preparation derived from each animals.

4. Assay principle

In FUJIFILM Wako Pure Chemical Corporation's LBIS Mouse Insulin ELISA Kit (H-type), biotinylated anti-insulin antibody, and standard or sample are incubated in monoclonal anti-insulin coated wells to capture insulin bound with biotinylated anti-insulin antibody. After 2 hours' incubation and washing, HRP (horse radish peroxidase)-conjugated streptavidin is added, and incubated for 30 minutes. After washing, HRP-conjugated streptavidin remaining in wells are reacted with a chromogen (TMB) for 30 minutes, and reaction is stopped by addition of acidic solution, and absorbance of yellow product is measured spectrophotometrically at 450 nm. The absorbance is proportional to insulin concentration. The standard curve is prepared by plotting absorbance against standard insulin concentrations. Insulin concentrations in unknown samples are determined using this standard curve.

5. Precautions

- For professional use only, beginners are advised to use this kit under the guidance of experienced person. In manual operation, proficiency in pipetting technique is recommended.
- Use clean laboratory glassware.

LBIS Mouse Insulin ELISA Kit (H-type)

- Avoid contact with the acidic stop solution and chromogen (TMB) containing hydrogen peroxide and tetramethylbenzidine. Wear gloves and eye and clothing protection when handling these reagents.
- Be careful not to allow the reagent solutions of the kit to touch the skin, eyes and mucus membranes. Especially be careful for the stop solution because it is sulfuric acid. The stop solution and the substrate solution may cause skin/eyes irritation. In case of contact with these wash skin/eyes thoroughly with water and seek medical attention, when necessary.
- Do not drink, eat or smoke in the areas where assays are carried out.
- In treating assay samples of animal origin, be careful for possible biohazards.
- This kit contains components of animal origin. These materials should be handled as potentially infectious.
- The materials must not be pipetted by mouth.
- Unused samples and used tips should be rinsed in 1 % formalin, 2 % glutaraldehyde, or more than 0.1 % sodium hypochlorite solution for more than 1 hour, or be treated by an autoclave before disposal.
- Dispose consumable materials and unused contents in accordance with applicable regional/national regulatory requirements.
- In order to avoid dryness of wells, contamination of foreign substances and evaporation of dispensed reagents, never forget to cover the well plate with a plate seal supplied, during incubation.
- ELISA can be easily affected by your laboratory environment. Room temperature should be at 20 °C - 25 °C strictly. Avoid airstream velocity over 0.4 m/sec. (including wind from air conditioner), and humidity less than 30 %.

6. Reagents supplied

Components	State	Amount
(A) Anti-Insulin coated plate	Use after washing	96 wells/1 plate
(B) Standard Insulin solution (200 ng/mL)	Concentrated. Use after dilution	300 µL/1 vial
(C) Buffer solution (bluish color)	Ready for use.	60 mL/1 bottle
(D) Biotinylated anti-insulin antibody	Concentrated. Use after dilution.	200 µL/1 vial
(E) HRP-conjugated streptavidin	Concentrated. Use after dilution.	200 µL/1 vial
(F) Chromogen (TMB)	Ready for use.	12 mL/1 bottle
(H) Stop solution	Be careful! Ready for use.	12 mL/1 bottle
(I) Wash stock solution (10×)	Concentrated. Use after dilution.	100 mL/1 bottle
Plate seal	—	3 sheets
Instruction Manual	—	1 copy

7. Equipments or supplies required but not supplied Use as a check box

- Deionized water (or Distilled water) Test tubes for preparation of standard solution series.
- Glassware for dilution of Wash stock solution (10×)(a graduated cylinder, a bottle). Pipettes (disposable tip type). One should be able to deliver 10 µL precisely, and another for 100µL - 500 µL. Syringe-type repeating dispenser like Eppendorf multipette plus which can dispense 100 µL. Paper towel to remove washing buffer remaining in wells. A vortex-type mixer. A shaker for 96 well-plate (600 rpm -1200 rpm). An automatic washer for 96 well-plate (if available), or a washing bottle with a jet nozzle. A 96 well-plate reader (450 nm ±10 nm, 620 nm: 600 nm - 650 nm) Software for data analysis.

8. Preparation of reagents

- ◆ Bring all reagents of the kit to room temperature (20 °C - 25 °C) before use.
- ◆ Prepare reagent solutions in appropriate volume for your assay. Do not store the diluted reagents.

【Concentrated reagents】

[(B) Standard Insulin solution (200 ng/mL)]

Make a serial dilution of master standard (200 ng/mL) solution to prepare each standard solution.

* Unit reduction for µIU/mL is 26 IU/mg. (Refer to 3. Introduction.)

LBIS Mouse Insulin ELISA Kit (H-type)

Volume of standard solution	(C) Buffer solution	Concentration (ng/mL)	Concentration (μ IU/mL)
Original solution: 100 μ L	100 μ L	100	2600
100 ng/mL solution: 100 μ L	100 μ L	50	1300
50 ng/mL solution: 100 μ L	100 μ L	25	650
25 ng/mL solution: 100 μ L	150 μ L	10	260
10 ng/mL solution: 100 μ L	100 μ L	5.0	130
5.0 ng/mL solution: 100 μ L	400 μ L	1.0	26
1.0 ng/mL solution: 100 μ L	100 μ L	0.5	13
0 (Blank)	100 μ L	0	0

[(D) Biotinylated anti-insulin antibody]

Prepare working solution by dilution of (D) with the (C) Buffer solution to **1:100**.

10 mL of the diluted solution is enough for 96 wells.

[(E)HRP-conjugated streptavidin]

Prepare working solution by dilution of (E) with the (C) Buffer solution to **1:100**.

10 mL of the diluted solution is enough for 96 wells.

[(I) Wash stock solution (10 \times)]

Dilute 1 volume of the concentrated Wash stock solution (10 \times) to **10 volume** with deionized water (or distilled water) to prepare working solution. Example: 100 mL of concentrated washing buffer (10 \times) and 900 mL of deionized water (or distilled water).

【Storage and stability】

[(A) Anti-Insulin coated plate]

If seal is not removed, put the strip back in a plastic bag with zip-seal originally used for well-plate container and store at 2 $^{\circ}$ C - 8 $^{\circ}$ C. The strip will be stable until expiration date.

[(B) Standard Insulin solution (200 ng/mL)]

Standard solutions prepared above should be used as soon as possible, and should not be stored.

[(C) Buffer solution] & [(F) Chromogen (TMB)]

If not opened, store at 2 $^{\circ}$ C - 8 $^{\circ}$ C. It maintains stability until expiration date. Once opened, we recommend using as soon as possible to avoid influence by environmental condition.

[(D) Biotinylated anti-insulin antibody] & [(E) HRP-conjugated streptavidin]

Unused working solution (already diluted) should be disposed.

[(H) Stop solution]

Close the stopper tightly and store at 2 $^{\circ}$ C - 8 $^{\circ}$ C. It maintains stability until expiration date.

[(I) Wash stock solution (10 \times)]

The rest of undiluted buffer: if stored tightly closed at 2 $^{\circ}$ C - 8 $^{\circ}$ C, it is stable until expiration date.

Dispose any unused diluted buffer.

9. Technical tips

- Be careful to avoid any contamination of assay samples and reagents. We recommend the use of disposal pipette tips, and 1 tip for 1 well.
- The reagents are prepared to give accurate results only when used in combination within the same box. Therefore, do not combine the reagents from kit with different lot numbers. Even if the lot number is the same, it is best not to mix the reagents with those that have been preserved for some period.
- Optimally, the reagent solutions of the kit should be used immediately after reconstitution. Otherwise, store them in a dark place at 2 $^{\circ}$ C - 8 $^{\circ}$ C.
- Time reaction from the pipetting of the reagent to the first well.
- Dilution of the assay sample must be carried out using the buffer solution provided in the kit.
- The substrate chromogen (TMB) should be almost clear pale blue before use. It turns blue during reaction, and gives yellowish color after addition of stop solution. Greenish color means incomplete mixing.
- To avoid denaturation of the coated antibody, do not let the plate go dry.
- As the anti-insulin-coated plate is module type of 8 wells \times 12 strips, each strip can be separated by cutting the cover sheet with a knife and used independently.
- When ELISA has to be done under the airstream velocity over 0.4 m/sec. and the humidity less than 30 %, seal the well plate with a plate seal and place the well plate in an incubator or a styrofoam box in each step

of incubation.

10. Preparation of samples

This kit is intended to measure insulin in mouse serum, plasma (preferably obtained with heparin), culture medium and tissue/cell extracts. The necessary sample volume for the standard procedure is 10 μ L. Samples should be immediately assayed or stored below $-35\text{ }^{\circ}\text{C}$ for several days. Defrosted samples should be mixed thoroughly for best results. Do not repeat freeze-and-thaw cycles.

Hemolytic and hyperlipemic serum samples are not suitable.

*** To avoid influence of blood (high lipid or hemolysis, etc.), if your original samples have heavy chyle or hemolysis, do not use them for assay. Abnormal value might be obtained with hemolysis above 40 mg/dL with this kit.**

If presence of interfering substance is suspected, examine by dilution test at more than 2 points. Dilution of a sample should be made in a test tube using buffer solution prior to adding them to wells. Turbid samples or those containing insoluble materials should be centrifuged before testing to remove any particulate matter.

Storage and stability

Insulin samples will be inactivated if stored at $2\text{ }^{\circ}\text{C} - 8\text{ }^{\circ}\text{C}$. If it is necessary to store sample in refrigerator ($2\text{ }^{\circ}\text{C} - 8\text{ }^{\circ}\text{C}$), add aprotinin at final concentration of 100 KIU/mL - 500 KIU/mL. (KIU: kallikrein inhibitor unit).

If you have to store assay samples for a longer period, snap-freeze samples and keep them below $-35\text{ }^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

11. Assay procedure

Remove the cover sheet of the anti-Insulin-coated plate after bringing up to room temperature.

- (1) Wash the anti-Insulin coated plate (A) by filling the wells with 300 μ L of washing buffer and discard 4 times (*②), then strike the plate upside-down onto several sheets of paper towel to remove residual buffer in the wells.
- (2) Pipette 100 μ L of biotinylated anti-insulin antibody to all wells. Shake the plate gently on a plate shaker (*③).
- (3) Pipette 10 μ L of sample to the designated sample wells.
- (4) Pipette 10 μ L of standard solution to the wells designated for standards.
- (5) Shake the plate gently on a plate shaker (*③).
- (6) Stick a plate seal (*④) on the plate and incubate for 2 hours at $20\text{ }^{\circ}\text{C} - 25\text{ }^{\circ}\text{C}$.
- (7) Discard the reaction mixture. Rinse wells by filling the wells with washing buffer and discard 4 times (*②), then strike the plate upside-down onto several sheets of paper towel to remove residual buffer in the wells.
- (8) Pipette 100 μ L of HRP-conjugated streptavidin to all wells, and shake as step (5).
- (9) Stick a plate seal (*④) on the plate and incubate the plate for 30 minutes at $20\text{ }^{\circ}\text{C} - 25\text{ }^{\circ}\text{C}$.
- (10) Discard the reaction mixture, and then wash the plate as step (1).
- (11) Pipette 100 μ L of chromogen (TMB) to wells, and shake as step (5).
- (12) Stick a plate seal (*④) on the plate and incubate the plate for 30 minutes at $20\text{ }^{\circ}\text{C} - 25\text{ }^{\circ}\text{C}$.
- (13) Add 100 μ L of the stop solution to all wells and shake as step (5).
- (14) Measure the absorbance of each well at 450 nm (reference wavelength, 620nm) immediately using a plate reader.

*Refer to the page 7 for notes of *②, *③ and *④.

12. Calculations


(1) Prepare a standard curve using semi-logarithmic or two-way logarithmic section paper by plotting absorbance* (Y-axis) against insulin concentration (ng/mL) on X-axis.

*Absorbance at 450 nm minus absorbance at 620 nm.

(2) Using the standard curve, read the insulin concentration of a sample at its absorbance*, and multiply the assay value by dilution factor if the sample has been diluted. Though the assay range is wide enough, in case the absorbance of some samples is higher than that of the highest standard, please repeat the assay after proper dilution of samples with the buffer solution.

* We recommend the use of 3rd order regression curve for log-log plot, or 4 or 5 parameters method for log-normal plot in computer calculation.

* Physiological or pathological situation of animals should be judged comprehensively taking other examination results into consideration.



Mouse insulin assay standard curve (an example)
Absorbance may change due to assay environment.

13. Performance characteristics

- Assay range of kit is 0.5 ng/mL - 100 ng/mL
- Specificity

The antibodies used in this kit are specific to insulin. Cross-reactivity of the kit is shown below.

Substances	Cross-reactivity	Remarks
Mouse C-peptide	—	At 10 ng/mL
Rat insulin	+	At 10 ng/mL
Human proinsulin	+	At 10 ng/mL

- Precision of assay
Within assay variation (3 samples, 5 replicates assay), Mean CV was within 10 %.
- Reproducibility
Between assay variation (3 samples, 3 days, assayed in duplicates), Mean CV was within 10 %
- Recovery test
Standard insulin was added in 3 concentrations to 2 serum samples and assayed.
The recoveries were 97.3 % - 106 %
- Dilution test
Two serum samples were serially diluted by 4 points.
The dilution curves showed excellent linearity with $R^2 = 0.9997 - 1.0$.

14. Reference assay data

Mean Insulin assay value: 1.59 ng/mL - 3.83 ng/mL, SD: 0.622 ng/mL - 2.11 ng/mL

Mouse strain: C57BL/6, KKAY, BALB/c, ICR, 7 w - 8 w males and females, ad libitum feeding

Number of animals: 8 - 12 / subspecies, serum or plasma

These data should be considered as guidance only. Each laboratory should establish its own normal and pathological reference ranges for insulin levels independently.

15. Trouble shooting

- Low absorbance in all wells
Possible explanations:
1) The standard or samples might not be added.
2) Reagents necessary for coloration such as biotinylated anti-insulin antibody, HRP-conjugated streptavidin, or chromogen (TMB) might not be added.
3) Wrong reagents related to coloration might have been added. Wrong dilution of biotinylated anti-insulin antibody or HRP-conjugated streptavidin.
4) Contamination of peroxidase enzyme inhibitor(s).
5) Influence of the temperature under which the kits had been stored.
6) Excessive hard washing of the well plate.
7) Addition of chromogen (TMB) soon after taking out from a refrigerator might cause poor coloration owing to low temperature.
- The OD of blank is higher than that of the lowest standard concentration (0.5 ng/mL)
Possible explanations:
Improper or inadequate washing. (Change washing repetition from 4 times to 5-8 times after the reaction with HRP-conjugated streptavidin.)
- High coefficient of variation (CV)
Possible explanation:
1) Improper or inadequate washing.
2) Improper mixing of standard or samples.
3) Pipetting at irregular intervals.
- Q-1: Can I divide the plate to use it for the other testing?
A-1: Yes, cut off the clear seal on the plate with cutter along strip. Put the residual plate, which is still the seal on, in a refrigerator soon
- Q-2: I found there contains liquid in 96 well-plate when I opened the box. What is it?
A-2: When we manufacture 96 well-plate, we insert preservation stabilizer in wells.

LBIS Mouse Insulin ELISA Kit (H-type)

Worksheet example


	Strip 1&2	Strip 3&4	Strip 5&6	Strip 7&8	Strip 9&10	Strip 11&12
A	100 ng/mL	Sample 1	Sample 9	Sample 17	Sample 25	Sample 33
B	50 ng/mL	Sample 2	Sample 10	Sample 18	Sample 26	Sample 34
C	25 ng/mL	Sample 3	Sample 11	Sample 19	Sample 27	Sample 35
D	10 ng/mL	Sample 4	Sample 12	Sample 20	Sample 28	Sample 36
E	5.0 ng/mL	Sample 5	Sample 13	Sample 21	Sample 29	Sample 37
F	1.0 ng/mL	Sample 6	Sample 14	Sample 22	Sample 30	Sample 38
G	0.5 ng/mL	Sample 7	Sample 15	Sample 23	Sample 31	Sample 39
H	0 (Blank)	Sample 8	Sample 16	Sample 24	Sample 32	Sample 40

LBIS Mouse Insulin ELISA Kit (H-type)

Summary of assay procedure : Use as a check box

*First, read this instruction manual carefully and start your assay after confirmation of details.

- Bring the well-plate and all reagents back to **20 °C - 25 °C for 2 hours.**
- Wash stock solution (10×) concentrate must be diluted to **10 times** by deionized water (or distilled water that returned to 20 °C - 25 °C).
- Standard Insulin solution dilution example:



- Biotinylated anti-insulin antibody(D) : Dilute to **100 times** by using the (C) Buffer solution and use.

<input type="checkbox"/> Anti-Insulin coated plate		
<input type="checkbox"/> ↓Washing 4 times (*②)		*⑥
<input type="checkbox"/> Biotinylated anti-insulin antibody	100 µL	
<input type="checkbox"/> ↓Shaking (*③)		
<input type="checkbox"/> Samples/Standards	10 µL	
<input type="checkbox"/> ↓Shaking (*③), Incubation for 2 hours at 20 °C - 25 °C. (Standing(*④))		
<input type="checkbox"/> HRP-conjugated streptavidin(E)		
<input type="checkbox"/> Dilute to 100 times by using (C) Buffer solution and use. (Dilute reagents during the first reaction.)		
<input type="checkbox"/> ↓Washing 4 times (*②)		*⑥
<input type="checkbox"/> HRP-conjugated streptavidin	100 µL	
<input type="checkbox"/> ↓Shaking (*③), Incubation for 30 minutes at 20 °C - 25 °C. (Standing(*④))		
<input type="checkbox"/> ↓Washing 4 times (*②)		*⑥
<input type="checkbox"/> Chromogen (TMB) (After dispense, the color turns to blue depending on the concentration.)	100 µL	
<input type="checkbox"/> ↓Shaking (*③), Incubation for 30 minutes at 20 °C - 25 °C. (Standing(*④))		
<input type="checkbox"/> Stop solution (After dispense, the color turns to yellow depending on the concentration.)	100 µL	
<input type="checkbox"/> ↓Shaking (*③) (Immediately shake.)		
<input type="checkbox"/> Measurement of absorbance (450 nm, Ref 620 nm) immediately (Ref. wave cancels the dirt in the back of plate.)		*⑤

*② After dispensing wash buffer to wells, lightly shake the plate on your palm for 10 seconds and remove the buffer. Guideline of washing volume: 300 µL/well for an automatic washer and for a pipette if the washing buffer is added by pipette. In case of washing by using 8 channel pipette, sometimes the back ground tends to be high. If so, change washing frequency from 4 times to 5-8 times at the constant stroke after the reaction with HRP-conjugated streptavidin. Standard of plate-washing pressure: 5 mL/min - 25 mL/min. (Adjust it depending on the nozzle's diameter.)

*③ Guideline of shaking: 600 rpm - 1200 rpm for 10 seconds × 3 times.


*④ Seal the plate during the reaction after shaking. Peel off the protective paper from the seal and stick the seal on the plate. Do not reuse the plate seal used once.

*⑤ 600 nm - 650 nm can be used as reference wavelength.

*⑥ After removal of wash buffer, immediately dispense the next reagent.

LBIS Mouse Insulin ELISA Kit (H-type)

Assay worksheet



LBIS Mouse Insulin ELISA Kit (H-type)

[Storage condition] Store the kit at 2 °C - 8 °C (Do not freeze).

[Term of validity] Expiration date is indicated on the container.

[Cat #] 630-10371

FUJIFILM Wako Pure Chemical Corporation

1-2, Doshomachi 3-Chome, Chuo-Ku, Osaka 540-8605, Japan

Telephone : +81-6-6203-3741

Facsimile : +81-6-6201-5964

<https://www.fujifilm.com/ffwk/en>

この度は弊社製品をご購入いただきましてありがとうございます。ご使用に際してはキットに同梱された取扱説明書に従って測定を実施してください。本キットを初めてご使用になられる場合は後述の「◆ご使用前にご確認頂きたい技術上のヒント及び注意事項」をご確認の上ご使用ください。

『 レビス® インスリン - マウス (Hタイプ) 』 取扱説明書

1. イントロダクション

インスリンは膵臓のランゲルハンス島（膵島）のβ細胞から分泌されるホルモンで、分子量は約5800、等電点5.4付近の蛋白質です。A6-A11、A7-B7、A20-B-19でS-S結合を形成し、酸性或いはZnの存在しない中性水溶液では2量体を形成しますが、中性でZn存在下ではZn2個を含む6量体を形成します。

肝、筋肉、脂肪組織が主要な標的組織ですが、それぞれに次のような作用を示します。

肝：グリコーゲン合成促進、蛋白合成促進、脂肪酸合成促進、糖利用の促進、糖新生抑制。

筋肉：糖、アミノ酸、Kの細胞膜透過性増大、グリコーゲン合成促進、蛋白合成促進、蛋白分解抑制。

脂肪組織：グルコースの細胞膜透過性増大、脂肪酸合成促進。

インスリンは細胞内で1本鎖のプロインスリンの形で合成された後、S-S結合が形成され、酵素分解による活性化がおこってC-ペプチドとインスリンが分離します。

*WHOはヒトインスリンの1st International Standard、1986として26 IU/mg (0.038 mg/IU)の精製品を提供しています。同時にウシインスリンについて1st International Standard、1986、25.7 IU/mg、ブタインスリン1st International Standard、1986、26 IU/mgを提供するようになりました。ヒトの場合、治療用に用いられますので、それに合わせてヒトの臨床検査での測定値もIUで表現する方が便利ですが、動物では重量で扱った方が良いでしょう。

上記のように精製インスリンはヒトで26 IU/mg、ウシで25.7 IU/mg、ブタで26 IU/mgとなっていますので、大体種を問わず26 IU/mg程度であると考えても良いでしょう。

本キットはマウスインスリンを定量的に測定するためのサンドイッチ酵素免疫測定法です。本キットは研究のみにご使用ください。

◆製品の特長

- 全反応時間は3時間です。
- マウス血清または血漿（ヘパリン血漿を推奨します）、培養上清、細胞抽出液中のインスリンを測定します。
- 微量な検体（標準操作法は10 μL）で測定可能です。
- 緩衝液に色（青色）がついており、分注済みウェルの確認が容易です。
- 1キットは96ウェルです。
- 全ての試薬は溶液タイプです。

2. 測定原理

本キットはビオチン結合抗インスリン抗体、標準品、検体を抗インスリンモノクローナル抗体固相化マイクロプレートウェル中で2時間インキュベートします。洗浄後、ペルオキシダーゼ・アビジン結合物を加え、捕捉されたインスリンとともに30分インキュベートします。再度の洗浄後、ウェルに残ったペルオキシダーゼを発色液（TMB）と反応させます。反応は酸性の溶液の添加で停止され、反応の結果生じた黄色の産物が450 nm（副波長620 nm）で比色測定されます。吸光度はインスリン濃度にほぼ比例します。標準品濃度に対して吸光度をプロットし標準曲線を作成し、この標準曲線から未知検体中の濃度が決定されます。

3. キットの保存と使用期限

キットは2℃～8℃で保存してください（凍結厳禁）。この保存条件下でキットは外箱のラベルに記載された有効期限内安定です。有効期限の過ぎた試薬は使用しないでください。開封した各試薬につきましては、保管状態により影響を受ける可能性がありますので早めのご使用を推奨します。

4. キット以外に必要な器具 □チェックリスト

- 精製水（蒸留水） □標準溶液希釈用試験管 □洗浄液希釈用ガラス器具（メスシリンダー・ビーカー・瓶） □チップ交換型ピペット（使い捨てチップで10 μLを正確にピペッティングできるもの、及び100 μL～500 μLを正確にピペッティングできるもの） □連続分注ピペット（例Eppendorfのmultipette plus）、100 μLを連続分注できるもの □ペーパータオル等の吸水性のあるもの（洗浄後にプレートに残った液を取り除く） □攪拌器（Vortexタイプ） □マイクロプレート振とう器（約600 rpm～1200 rpm） □96ウェルプレート用洗浄機（あれば好ましい）または噴射ビン □96ウェルプレートリーダー（450 ± 10 nm、620 nm：600 nm～650 nm） □データ計算用ソフトウェア

5. 構成成品

構成成品	状態	容量
(A) Anti-Insulin coated plate 抗体固相化 96 ウェルプレート	洗浄後使用	96 wells(8×12)/1 枚
(B) Standard Insulin solution (200 ng/mL) 標準インスリン溶液 (200 ng/mL)	希釈後使用	300 µL/1 本
(C) Buffer solution (bluish color) 緩衝液 (青色)	そのまま使用	60 mL/1 本
(D) Biotinylated anti-insulin antibody ビオチン結合抗インスリン抗体	希釈後使用	200 µL/1 本
(E) HRP-conjugated streptavidin ペルオキシダーゼ・アビジン結合物	希釈後使用	200 µL/1 本
(F) Chromogen (TMB) 発色液(TMB)	そのまま使用	12 mL/1 本
(H) Stop solution 反応停止液	Be careful! 取扱注意	そのまま使用 12 mL/1 本
(I) Wash stock solution (10×) 濃縮洗浄液(10×)	希釈後使用	100 mL/1 本
Plate seal プレートシール		3 枚
Instruction Manual 取扱説明書		1 部

6. 試薬の調製

- * キットの試薬は使用前に必ず室温(20 °C~25 °C)に戻してください (2 時間位が目安です)。
- * 5 で「そのまま使用」とある試薬は室温化後そのままの状態で使用できます。「希釈後使用」とあるものについては下記の要領で調製してください。
- * 測定に必要な分だけ試薬を調製してください (ご不明な際にはお問い合わせください)。

【濃縮された試薬類】

[(B) Standard Insulin solution (200 ng/mL)] ; 標準曲線作成用

(B) Standard Insulin solution (200 ng/mL) (原液) と(C) Buffer solution (bluish color)を使って標準溶液を調製してください。下記は一例です。※µIU/mL 換算は 26 IU/mg で行っております (イントロダクション参照)

標準溶液の容量	(C) Buffer solution (bluish color)	濃度 (ng/mL)	濃度 (µIU/mL※)
標準溶液原液 100 µL	100 µL	100	2600
100 ng/mL 溶液 100 µL	100 µL	50	1300
50 ng/mL 溶液 100 µL	100 µL	25	650
25 ng/mL 溶液 100 µL	150 µL	10	260
10 ng/mL 溶液 100 µL	100 µL	5.0	130
5.0 ng/mL 溶液 100 µL	400 µL	1.0	26
1.0 ng/mL 溶液 100 µL	100 µL	0.5	13
0(Blank)	100 µL	0	0

[(D) Biotinylated anti-insulin antibody]

200 µL を充分分取できる量をご提供しています。濃縮液を(C) Buffer solution (bluish color)で **100 倍**に希釈してください。

[(E) HRP-conjugated streptavidin]

200 µL を充分分取できる量をご提供しています。濃縮液を(C) Buffer solution (bluish color)で **100 倍**に希釈してください。

[(I) Wash stock solution (10×)]

濃縮洗浄液(10×)を室温化された精製水 (蒸留水) で **10 倍**に希釈してください。

例 : 100 mL の濃縮洗浄液(10×)+ 900 mL の精製水 (蒸留水) (96 ウェル全てを使用する場合)

【試薬の安定性と保存方法】

(A) Anti-Insulin coated plate

未使用（冷蔵状態を保った状態でシールを剥がしていない）抗体固相化ストリップは同梱のジップシールパックに戻し、そのまま 2℃～8℃で保存してください。有効期限内安定性を保ちます。

(B) Standard Insulin solution (200 ng/mL)

キットを分割して使用する際は使用する直前に冷蔵庫より取り出し希釈調製し、残りの原液は室温に戻さないで直ちに蓋をしっかりと閉め、2℃～8℃で保存してください。有効期限内安定性を保ちます。希釈した各標準溶液は直ちに使用し、保存はしないでください。

(C) Buffer solution (bluish color)及び(F) Chromogen (TMB)

一部の溶液を使用する際は必要量より少し多めの量を別の容器に移し、残りは室温に戻さないで直ちに蓋をしっかりと閉め、2℃～8℃で保存してください。有効期限内安定性を保ちます。

(D) Biotinylated anti-insulin antibody 及び(E) HRP-conjugated streptavidin

キットを分割して使用する際は希釈時に冷蔵庫より取り出し希釈調製し、残りの原液は室温に戻さないで直ちに蓋をしっかりと閉め、2℃～8℃で保存してください。有効期限内安定性を保ちます。使用残りの希釈済み液は廃棄してください。

(H) Stop solution

使用残りを保存する場合は、蓋をしっかりと閉め、2℃～8℃で保存してください。有効期限内安定性を保ちます。

(I) Wash stock solution (10×)

濃縮洗浄液(10×)を保存する場合は、蓋をしっかりと閉め、2℃～8℃で保存してください。有効期限内安定性を保ちます。使用残りの希釈済み洗浄液は廃棄してください。

7.検体の調製

本キットはマウス血清または血漿（ヘパリン血漿を推奨します）、培養上清、細胞抽出液中のインスリンを測定します。

- 検体は採取後すぐに測定するか、-35℃以下で凍結保存してください。凍結した検体は測定する直前に解凍し十分に攪拌してください。繰り返しの凍結融解は避けてください。正しい結果が得られない原因になります。
- 溶血した検体や高脂質検体は使わないでください。
※血液成分の影響（高脂質・溶血等）を抑制する為に原検体中の脂質（乳ビ）・溶血が高い場合は異常値発生の原因となる場合がありますので測定に使用しないでください。本キットの場合、溶血は 40 mg/dL 以上で影響が現れます。
- 濁り及び不溶物のある検体は遠心分離等で除去後測定に用いてください。
- 妨害物質の影響が疑わしい検体は、同一検体において、異なる 2 ポイント以上の希釈率で希釈直線性を確認してください。
- 検体を希釈する場合は、あらかじめ試験管等を用いて緩衝液で希釈し測定ウェルに分注してください。

【検体の安定性と保存方法】

インスリン分解酵素等のプロテアーゼの働きを抑えるため、採血時に最終濃度が 100 KIU/mL～500 KIU/mL のアプロチニンを添加して保管することをお勧めします。また、長期に保管する場合は、-35℃以下での凍結保管を推奨します。繰り返しの凍結融解は避けてください。(KIU : kallikrein inhibitor unit)

8.測定操作法

洗浄操作を始める前に次に分注する試薬を前もって用意してください。

抗体固相化プレートのシールは、プレートが十分に室温に戻ってから剥がしてください。

- (1) あらかじめ調製した洗浄液を各ウェルに満たし 4 回洗浄(*①)します。その後ペーパータオルなどの上でプレートを逆さにし、軽く叩きつけるようにしてウェルに残った液を取り除きます。
- (2) 各ウェルにビオチン結合抗インスリン抗体を 100 µL ずつ分注します。マイクロプレート振とう器などを用いて攪拌(*②)します。
- (3) 検体測定ウェルに検体を 10 µL 添加します。
- (4) 標準品測定ウェルに各濃度の標準溶液を 10 µL ずつ分注します。
- (5) マイクロプレート振とう器などを用いて攪拌(*②)します。
- (6) プレートシールを貼り、室温(20℃～25℃)で 2 時間静置(*③)します。
- (7) 反応終了後、反応液を捨て洗浄液を各ウェルに満たし、4 回洗浄(*①)します。その後、ペーパータオルなどの上でプレートを逆さにし、軽く叩きつけるようにしてウェルに残った液を取り除きます。
- (8) 各ウェルにペルオキシダーゼ・アビジン結合物を 100 µL ずつ分注します。マイクロプレート振とう器などを用いて攪拌(*②)します。
- (9) プレートシールを貼り、室温(20℃～25℃)で 30 分間静置(*③)します。
- (10) 反応終了後、反応液を捨て洗浄液を各ウェルに満たし 4 回洗浄(*①)します。その後、ペーパータオルなどの上でプレートを逆さにし、軽く叩きつけるようにしてウェルに残った液を取り除きます。
- (11) 各ウェルに発色液を 100 µL ずつ分注します。マイクロプレート振とう器などを用いて攪拌(*②)します。
- (12) プレートシールを貼り、室温(20℃～25℃)で 30 分間静置(*③)します。
- (13) 各ウェルに反応停止液を 100 µL ずつ分注し、発色反応を停止します。
- (14) 攪拌(*②)後、直ちにマイクロプレート用分光光度計で 450 nm（副波長 620 nm）での吸光度を測定します。副波長は 600 nm～650 nm の範囲で使用できます。

(*①)、(*②)、(*③)測定手順概要 (14、15 ページ) をご参照ください。


9.計算

(1)測定毎に標準曲線を作成します。両対数を使用し X 軸を標準溶液濃度(ng/mL)、Y 軸を吸光度の標準曲線グラフを作成してください。

(2)標準曲線より、検体の吸光度に対応する濃度 (ng/mL)を読み取ります。

* 検体の吸光度が標準曲線吸光度より外れた場合は(C) Buffer solution (bluish color)にて適当倍率に調製し再度測定を実施してください。

* 演算処理では、3 次多項式または 4 または 5 パラメーターの使用をお勧め致します。



グラフは標準曲線例です (吸光度は、測定環境により変動します)。

* プレートリーダーは SUNRISE RAINBOW(TECAN)を使用

10.キットの性能

●測定範囲

マウスインスリンを 0.5 ng/mL~100 ng/mL の範囲で測定できます。

●特異性

関連物質を本キットで測定した結果は次表のとおりです。

※交差性は、10 ng/mL 濃度時のデータです。

+ : 交差有り / - : 交差無し / N.T. : Not tested

検体名	交差性	検体名	交差性
マウス C-ペプチド	-	ラット インスリン	+
マウス プロインスリン	N.T.	ヒト インスリン	+

●精度試験 (アッセイ内変動) (5 重測定、3 検体)

平均 C.V.値は 10 %未満

●再現性試験 (アッセイ間変動) (2 重測定、3 検体、3 日間)

平均 C.V.値は 10 %未満

●添加回収試験

2 血清検体に異なる 3 濃度のインスリンを添加し測定した結果、回収率は 97.3 %から 106 %

●希釈直線性

2 血清検体を連続的に希釈用緩衝液で 4 段階希釈し測定した結果、直線回帰の R² は 0.9997 と 1

11.参考値

マウスインスリン測定値 : 平均 1.59 ng/mL~3.83 ng/mL、標準偏差 0.622 ng/mL~2.11 ng/mL

亜種 : C57BL/6、KKAy、BALB/c、ICR、雌雄、7~8 週齢、不断給餌、8~12 匹/亜種、血清/血漿飼育条件、採血条件、検体保管条件により測定値は変動しますので、この測定値は目安としてお使いください。

12.トラブルシューティングと Q&A

●すべてのウェルでの反応が弱い

原因として考えられること

- 1)標準品や検体の入れ忘れ。
- 2)発色に関連する試薬溶液の入れ忘れ。
- 3)発色に関連する試薬溶液の取り違いや希釈調製不良。
- 4)ペルオキシダーゼ酵素阻害剤の混入。


LBIS Mouse Insulin ELISA Kit (H-type)

- 5)キット保管温度の影響（凍結した場合）。
- 6)プレートの過剰な洗浄。
- 7)発色液の温度が低かった。
- 最小標準溶液濃度(0.5 ng/mL)の OD 値よりブランク OD 値が高くなる
原因として考えられること
洗浄が不適當、不完全であった。
(ペルオキシダーゼ・アビジン結合物と反応後の洗浄回数 4 回を同じ流速で 5 回~8 回に増やしてください。)
- 変動係数(CV)が大きい
原因として考えられること
1)洗浄が不適當、不完全であった。
2)標準品や管理血清、または検体の攪拌が不充分であった（凍結検体の攪拌は充分に行ってください）。
3)ピペティング操作が一定ではなかった。
- Q-1：キットは分割して使用することができますか？
A-1：できます。プレートに貼られた透明シールをストリップの間にそってカッターなどで切り離してご使用ください。使用しないプレートはシールを貼った状態で冷蔵庫に保管してください。
- Q-2：プレートを取出したらウェルの中に液体が入っていましたが何ですか？
A-2：出荷時には保存安定液が充填してあります。

【測定手順概要とチェックリスト】

必ず取扱説明書を一読して検体条件、測定条件、測定方法を確認後測定操作を行ってください。

- ウェルプレート、試薬類を十分に室温(20℃～25℃)に戻してください。**室温化には2時間位必要**
- 濃縮洗浄液の希釈：室温化された精製水で、**10倍**に希釈してください。
- 標準溶液の希釈(例)：室温化された緩衝液で、希釈してください。



- ビオチン結合抗インスリン抗体の希釈：室温化された緩衝液で **100倍**に希釈してください。

各操作注意事項並びに関連情報

<input type="checkbox"/>	抗体固相化 96 ウェルプレート		
<input type="checkbox"/>	↓洗浄4回 (洗浄液除去後、直ちに次の試薬分注)		*①
<input type="checkbox"/>	ビオチン結合抗インスリン抗体	100 µL	
<input type="checkbox"/>	↓攪拌		*②
<input type="checkbox"/>	検体 または 標準インスリン溶液	10 µL	
<input type="checkbox"/>	↓攪拌、室温 (20℃～25℃)、2時間反応、静置		*② *③
<input type="checkbox"/>	ペルオキシダーゼ・アビジン結合物の希釈。室温化された緩衝液で、 100倍 に希釈してください。希釈溶液の調製は第一反応中に行う。		
<input type="checkbox"/>	↓洗浄4回 (洗浄液除去後、直ちに次の試薬分注)		*①
<input type="checkbox"/>	ペルオキシダーゼ・アビジン結合物	100 µL	
<input type="checkbox"/>	↓攪拌、室温 (20℃～25℃)、30分間反応、静置		*② *③
<input type="checkbox"/>	↓洗浄4回 (洗浄液除去後、直ちに次の試薬分注)		*①
<input type="checkbox"/>	発色液(TMB) 分注後、濃度により青色に変色	TMBが室温化されていることを確認	100 µL
<input type="checkbox"/>	↓攪拌、室温 (20℃～25℃)、30分間反応、静置		*② *③
<input type="checkbox"/>	反応停止液 分注後、濃度により黄褐色に変色	強酸性につき取扱注意	100 µL
<input type="checkbox"/>	↓攪拌 (直ちに攪拌)		*②
<input type="checkbox"/>	直ちに吸光度測定 (主波長 450 nm、副波長 620 nm:600 nm～650 nm) 副波長はプレート裏面の汚れ等をキャンセルします		

(*①) 洗浄液をウェルに分注後、手のひらの上で10秒ほど軽く振り廃棄します。4回連続洗浄後、ペーパータオル上にプレートを逆さにして叩き洗浄液を完全に除去します。洗浄液除去後の乾燥に注意して次の溶液を直ちに分注します。洗浄液をピペットで添加する際の液量目安は300 µL/ウェルです。プレート洗浄機ご使用の場合の圧力目安は5 mL/分～25 mL/分 (ノズルの径により異なります) です。第一反応後の初回の洗浄のみウェル間のコンタミに注意してください。

(*②) 攪拌の目安は600 rpm～1200 rpm-10秒間、3回。

(*③) 攪拌終了後プレートシールを貼り静置してください。

プレートシールは保護紙を剥がして、粘着面をプレート側にして貼り付けてください。一度使用した

プレートシールは再使用しないでください。


ワークシート (例)

	Strip 1&2	Strip 3&4	Strip 5&6	Strip 7&8	Strip 9&10	Strip 11&12
A	100 ng/mL	検体 1	検体 9	検体 17	検体 25	検体 33
B	50 ng/mL	検体 2	検体 10	検体 18	検体 26	検体 34
C	25 ng/mL	検体 3	検体 11	検体 19	検体 27	検体 35
D	10 ng/mL	検体 4	検体 12	検体 20	検体 28	検体 36
E	5.0 ng/mL	検体 5	検体 13	検体 21	検体 29	検体 37
F	1.0 ng/mL	検体 6	検体 14	検体 22	検体 30	検体 38
G	0.5 ng/mL	検体 7	検体 15	検体 23	検体 31	検体 39
H	0 (Blank)	検体 8	検体 16	検体 24	検体 32	検体 40

◆ご使用前にご確認頂きたい技術上のヒント及び注意事項

- インスリン分解酵素等のプロテアーゼの働きを抑えるため、採血時に最終濃度が 100 KIU/mL～500 KIU/mL のアプロチニンを添加して保管することをお勧めします。また、長期に保管する場合は、-35℃以下での凍結保管を推奨します。繰り返しの凍結融解は避けてください。(KIU : kallikrein inhibitor unit)
- ELISA 法は測定環境により影響を受けます。測定操作、静置反応場所の室温：20℃～25℃(実験台上またはインキュベータ内温度)を厳守してください。また、風速(エアコンの風も含む)：0.4 m/sec 以上、湿度 30%未満の環境下での測定は避けてください。やむを得ず、測定操作を風速：0.4 m/sec 以上、湿度 30%未満の環境下で実施する場合には、各ステップの静置反応時、プレートシールをすることに加え、下記のような方法をご検討ください。
例) インキュベータ内、発泡スチロール製箱内で静置反応させる等。測定室の環境条件により対策方法が異なる場合がありますので、詳細はお問い合わせください。
- 各ステップでの静置反応時には、ウェルの乾燥、異物の混入、温度の偏り、分注試薬の蒸発を防止する為、必ずプレートシールを貼ってください。
- 検体と試薬に不純物が混ざらないように気をつけてください。1 ウェル/1 チップのご使用をお勧めします。
- 発色液は 96 ウェルプレートに使用するまでは薄い青色澄明です。光を避けて保存してください。
- 反応停止液は使用するまでは無色です。
- 本キットは ELISA 法の研修を終了した方、または指導者の方でご使用ください。的手法操作で測定する際にはピペティング操作の再現性が安定した方がご使用ください。
- 準備並びに本キット操作中は手袋、眼鏡、保護用着衣を身につけてください。
- 試薬類を皮膚に付けないでください。本キットの試薬が誤って、目、口、傷口、皮膚等に付着した場合は直ちに水道水で十分に洗い流す等の応急処置を行い、必要な場合は医師の手当てを受けてください。
- 本キットを使用している場所では飲食や喫煙をしないでください。
- 試薬類は口でピペティングしないでください。
- ロット番号の違う試薬とは混ぜて使わないでください。
- 検体は感染の危険性があるものとして充分注意して取り扱ってください。本キットは動物由来の成分を含んでいます。
- 使用済みの検体、使用した消耗品等は 1%ホルマリン、2%グルタールアルデヒドまたは 0.1%以上の次亜塩素酸ナトリウム溶液に 1 時間以上浸けてください。またはオートクレーブ滅菌処理して廃棄してください。使用した消耗品や未使用の薬品類は所属施設の規定並びに各地域の法令に従って廃棄してください。

LBIS Mouse Insulin ELISA Kit (H-type)



【測定名】

【所属】

【測定者】

【測定日】

【ロット番号】

【有効期限】

【備考】

【製品名】

レビス® インスリン-マウス (Hタイプ)

【和光コード】

630-10371

【英語表記】

LBIS Mouse Insulin ELISA Kit (H-type)

【お問い合わせ先】

製造発売元

富士フイルム 和光純薬株式会社

大阪府中央区道修町三丁目 1 番 2 号

Tel : 06-6203-3741

Fax : 06-6201-5964

<https://www.fujifilm.com/ffwk/ja>