

FUJIFILM**Wako**Code No. 013-28154 (5 mL)
019-28151 (1 mL×5A)

for Amino Acid Automated Analysis
**Amino Acids Mixture Standard Solution,
Type H (High Range)**
アミノ酸混合標準液, H 型 (高濃度タイプ)

Contents :

Amino Acids	Specification (μmol/mL)
L-Aspartic Acid	2.50 (2.38 ~ 2.62)
L-Threonine	2.50 (2.38 ~ 2.62)
L-Serine	2.50 (2.38 ~ 2.62)
L-Glutamic Acid	2.50 (2.38 ~ 2.62)
Glycine	2.50 (2.38 ~ 2.62)
L-Alanine	2.50 (2.38 ~ 2.62)
L-Valine	2.50 (2.38 ~ 2.62)
L-Cystine	2.50 (2.38 ~ 2.62)
L-Methionine	2.50 (2.38 ~ 2.62)
L-Isoleucine	2.50 (2.38 ~ 2.62)
L-Leucine	2.50 (2.38 ~ 2.62)
L-Tyrosine	2.50 (2.38 ~ 2.62)
L-Phenylalanine	2.50 (2.38 ~ 2.62)
Ammonium Chloride	2.50 (2.38 ~ 2.62)
L-Lysine	2.50 (2.38 ~ 2.62)
L-Histidine	2.50 (2.38 ~ 2.62)
L-Arginine	2.50 (2.38 ~ 2.62)
L-Proline	2.50 (2.38 ~ 2.62)

* Store in the dark at 2~10°C *

※各ロットの調液値は別途成績書にてご確認下さい。
※Please check the concentration of each lot in the report.

Notes :

1. This standard solution contains approximately 1.2 mL of solution of 0.1 mol/L HCl.

FUJIFILM Wako Pure Chemical Corporation

1-2, Doshomachi 3-Chome, Chuo-Ku, Osaka 540-8605, Japan
Telephone : + 81-6-6203-3741
Facsimile : + 81-6-6201-5964
<http://flwk.fujifilm.co.jp>

FUJIFILM Wako Chemicals U.S.A. Corporation

1600 Bellwood Road
Richmond, VA 23237
U.S.A.
Telephone : + 1-804-271-7677
Facsimile : + 1-804-271-7791
<http://www.wakousa.com>

FUJIFILM Wako Chemicals Europe GmbH

Fuggerstrasse 12
D-41468 Neuss
Germany
Telephone : + 49-2131-311-0
Facsimile : + 49-2131-311100
<http://www.wako-chemicals.de>

2304KA4