

FUJIFILM**Wako**

For Research Use Only. Not for use in diagnostic procedures.

Code No. 125-07021 (80 μ L)**LBIS™ Dog Insulin Standard Solution****Volume**80 μ L/bottle**Insulin Concentration**

240 ng/mL

Purification Method

Reversed-phase chromatography

Buffer

Phosphate buffer (with stabilizer)

Storage

2°C - 10°C

Procedure

Dog insulin can be quantitatively measured by replacing the Insulin Standard included in the LBIS™ Rat Insulin ELISA Kit (Code No. 299-94801) with this product. When using a portion of this product, remove it from the refrigerator immediately before use to prepare the dilution. Immediately and tightly close the remaining stock solution without allowing it to return to room temperature and store it at 2°C - 10°C. Use the prepared dilutions promptly and do not store them.

Dog Insulin Standard Solution (240 ng/mL) for Standard Curve Preparation

Prepare standard solutions using the LBIS™ Dog Insulin Standard Solution (240 ng/mL, stock solution) and (C) Buffer supplied with the kit. The following is an example. *Conversion to μ IU/mL is calculated using 26 IU/mg.

Standard Solution Volume	(C) Buffer	Concentration (ng/mL)	Concentration (μ IU/mL*)
10 μ L of Standard stock solution	190 μ L	12	312
100 μ L of 12 ng/mL solution	100 μ L	6.0	156
100 μ L of 6.0 ng/mL solution	100 μ L	3.0	78
100 μ L of 3.0 ng/mL solution	100 μ L	1.5	39
100 μ L of 1.5 ng/mL solution	100 μ L	0.75	19.5
100 μ L of 0.75 ng/mL solution	100 μ L	0.375	9.75
100 μ L of 0.375 ng/mL solution	100 μ L	0.188	4.88
0 (Blank)	100 μ L	0	0

Expiration Date

6 months from the date of manufacture (when unopened and kept refrigerated). The expiration date is indicated on the outer label of the kit.

Precaution

As the Standard Solution may be dispersed inside the bottle, collect it at the bottom by centrifugation or similar methods before opening the cap.

(Recommended centrifugation conditions : 500 g - 1,000 g, 10 second - 15 seconds, 2°C - 10°C)

LBIS™ Dog Insulin Standard Solution

[Storage] Store at 2°C - 10°C

[Expiration Date] Indicated on the label.

[Package] 80 μ L

[Cat #] 125-07021

FUJIFILM Wako Pure Chemical Corporation

1-2, Doshomachi 3-Chome, Chuo-Ku, Osaka 540-8605, Japan
Telephone : + 81-6-6203-3741
Facsimile : + 81-6-6201-5964
<http://fwk.fujifilm.co.jp>

Group Companies



Distributors



本品は研究用試薬です。体外診断用として使用できません。

Code No. 125-07021 (80 μ L)

レビス™ イヌインスリン標準液

容量

80 μ L/ 本

インスリン濃度

240ng/mL

精製法

逆相クロマトグラフィー

緩衝液

リン酸緩衝液（安定化剤含有）

保管

2～10℃

使用方法

レビス™ ラットインスリン ELISA キット（Code No. 299-94801）付属のインスリン標準品を本品に変更することによりイヌインスリンを定量的に測定できます。分割して使用する際は使用する直前に冷蔵庫より取り出し希釈調製し、残りの原液は室温に戻さないで直ちに蓋をしっかりと閉め、2℃～10℃で保存して下さい。希釈した各標準溶液は直ちに使用し、保存はしないで下さい。

イヌインスリン標準溶液（240ng/mL）；標準曲線作成用

レビス™ イヌインスリン標準液（240ng/mL）（原液）とキット添付（C）緩衝液を使って標準溶液を調製して下さい。下記は一例です。※ μ IU/mL 換算は 26 IU/mg で行っております。

標準溶液の容量	(C) 緩衝液	濃度 (ng/mL)	濃度 (μ IU/mL ※)
標準溶液原液 10 μ L	190 μ L	12	312
12ng/mL 溶液 100 μ L	100 μ L	6.0	156
6.0ng/mL 溶液 100 μ L	100 μ L	3.0	78
3.0ng/mL 溶液 100 μ L	100 μ L	1.5	39
1.5ng/mL 溶液 100 μ L	100 μ L	0.75	19.5
0.75ng/mL 溶液 100 μ L	100 μ L	0.375	9.75
0.375ng/mL 溶液 100 μ L	100 μ L	0.188	4.88
0 (Blank)	100 μ L	0	0

使用期限

製造後 6 か月：使用期限はキット外箱のラベルに記載（未開封、未融解時）

注意

標準液が容器内に分散している場合がありますので、キャップを開ける前に遠心分離等により標準液を容器の底に集めご利用下さい。

（遠心分離条件例：500～1000g、10 秒～15 秒、2～10℃）

【製品名】	レビス™ イヌインスリン標準液
【和光コード】	125-07021
【英語表記】	LBIS™ Dog Insulin Standard Solution
【貯法】	2～10℃ 保存
【使用期限】	ラベルに記載
【包装】	80 μ L/ 本×1 本

製造発売元

富士フイルム 和光純薬株式会社

大阪市中央区道修町三丁目1番2号

Tel：06-6203-3741