

## SPECIFICATION

## Acetonitrile for Pesticide Residue And Polychlorinated Biphenyl Analysis

Concentration 5000

REQUIREMENT	SPECIFICATION
* Appearance	Colorless clear liquid
** Assay (CH <sub>3</sub> CN) (GC)	min.99.8%(mass/mass)
Density (20°C)	0.780~0.783g/mL
Water	max.0.1%(mass/mass)
Residue after evaporation	max.0.0002%(mass/mass)
Suitability for pesticide residue-PCB analysis	to pass test
* Suitability for pesticide residue analysis (GC-FPD)	to pass test

JIS K 8039:2012(\*:Additional test performed by Wako but not required by JIS)(\*\*:The specification is originally set by Wako not JIS)