

Code No.03-2495

## SPECIFICATION

Chlorate Ion Standard Solution (ClO<sub>3</sub><sup>-</sup> 1000)

REQUIREMENT	SPECIFICATION
Appearance	Colorless clear liquid
Foreign anions	to pass test
Concentration (20°C)	980~1020mg/L