SPECIFICATION

Germanium Standard Solution (Ge 1000)

for ICP Analysis

REQUIREMENT	SPECIFICATION
Appearance	Colorless clear liquid
Acidity (as HNO3)	4∼6w/w%
Concentration (20℃)	990~1010mg/L
Trace impurities	to pass test

(1 / 1) revised on 2020/10/06