

SPECIFICATION

Lithium Tetraborate, Granules
for X-ray Fluorometry

REQUIREMENT	SPECIFICATION
Appearance	White, shot ~ granules
Loss on ignition	max.1.0%
Chloride (Cl)	max.0.005%
Sulfate (SO ₄)	max.0.005%
Sodium (Na)	max.0.02%
Magnesium (Mg)	max.0.001%
Calcium (Ca)	max.0.001%
Aluminium (Al)	max.0.002%
Silicon (Si)	max.0.005%
Titanium (Ti)	max.0.005%
Manganese (Mn)	max.0.001%
Iron (Fe)	max.0.002%
Assay	98.0~102.0%