

Code No.12-0532

SPECIFICATION

Ligroin

Guaranteed Reagent

REQUIREMENT	SPECIFICATION
* Appearance Boiling range (80~110°C) Residue after evaporation Acidity (as H ₂ SO ₄) Sulfide and reducing substances Substances darkened by sulfuric acid Copper corrosion test	Colorless clear liquid min.90%(volume/volume) max.0.001%(mass/mass) max.5ppm(mass/mass) to pass test to pass test to pass test

JIS K 8937:2020(*:Additional test performed by Wako but not required by JIS)