2024/10/12
Distributed by FUJIFILM Wako Pure Chemical Corporation
Manufactured by FUJIFILM Wako Pure Chemical Corporation

Code No.19-1379

SPECIFICATION

Styrene, Monomer Standard Solution (1mg/mL Methanol Solution)

for Water Analysis

REQUIREMENT	SPECIFICATION
Appearance	Colorless clear liquid
Practical test	to pass test
Concentration	0.97～1.03mg/mL

revised on 2022/04/28