

Code No.20-1300

SPECIFICATION

0.2mol/L Tris-HCl Buffer Solution for Endotoxin Detection

REQUIREMENT	SPECIFICATION
Appearance	Colorless clear liquid
pH at 25°C	7.9~8.1
Bacterial endotoxins	to pass test

Manufacturer FUJIFILM Healthcare Manufacturing Corporation