

Code No.29-3195

# SPECIFICATION

## Presep(R)-C Florisil(R) for Sample Pretreatment

REQUIREMENT	SPECIFICATION
Appearance	to pass test
Loss on drying at 180°C	max.3.0%
Weight of column packing	under investigation
Permitted pressure (1)	to pass test
Permitted pressure (2)	to pass test
Leakage of column packing	to pass test
Recovery	95~105%
State of column packing	under investigation