

Code No.29-3409

SPECIFICATION

PresepR C18(ODS) (2g/15mL) for Sample Pretreatment

REQUIREMENT	SPECIFICATION
Appearance	to pass test
Loss on drying at 85°C	max.5.0%
Weight of column packing	1.9~2.1g
Leakage of column packing	to pass test
Recovery	90~105%